Characterization, 'The Turbulator'. PID signals measured using Aurora miniPID

Preferred operating parameters:

central carrier flow:
3 L/min

odorant flow rate:
50 - 100 ml/min

(50 mL/min recommended; 40 mL/min OK).

Nozzle position approx.
5 cm from nose

Data courtesy of Wachowiak lab, Shawn Burton, 2018